



Order form

Date:

Client's data: _____

Rf.user*: _____

*You are required **NOT** to indicate personal and sensitive data of the user in accordance with General Data Protection Regulation EU 2016/679.

CONFIGURATION

Mobile stander that assists the user during vertical position and permits mobility exercise

01	Model	<input type="checkbox"/> JUPITER Standard	<input type="checkbox"/> JUPITER Big
		02 Basic configuration	
		24" wheel <input checked="" type="checkbox"/>	<input checked="" type="checkbox"/> 32" wheel
		<ul style="list-style-type: none"> ✓ 3 swivel castors with individual brakes ✓ chest support with soft padding ✓ adjustable knee supports ✓ individual footrests ✓ tray 	
03	Accessories	<input type="checkbox"/> JUP-KIT-01	Couple of black aluminium pushrims 24"
		<input type="checkbox"/> JUP-KIT-04	Couple of black aluminium pushrims 32"
		<input type="checkbox"/> JUP-KIT-02	Couple of main wheel regulation blocks
		<input type="checkbox"/> JUP-KIT-03	Panty
04	NOTES	For better device configuration, please indicate weight, height and age of the user. Describe the user indicating diagnosis, posture characteristics or other relevant details.	
		<p>① User's we kg _____</p> <p>② User's hei cm _____</p> <p>③ User's agi yrs _____</p>	<i>(filled by healthcare professionals)</i>

signature _____

TECHNICAL DATA		STANDARD		BIG	
Minimum user's height	cm	75		100	
Maximum user's weight	cm	110		140	
Peso dispositivo	kg	14		18	
Tyre pressure	bar	6-8		N/A	
Portata massima dispositivo	kg	40			
Rear wheels chambered position	°	4			
Rear wheels diameter	"	24"		32"	
Total length	mm	770		920	
Total width	mm	580		690	
Total height	mm	max 920	max 760	max 1100	min 960
Footrest-chest support distance	mm	max 880	min 650	max 1050	min 880
Chest support-knee support distance	mm	max 630	min 310	max 750	min 400
Knee support-footrest distance	mm	max 540	min 210	max 605	min 330
Knee cups span	mm	max 230	min 130	max 260	min 130
Individual footrests span	mm	max 230	min 110	max 310	min 110
Footrest height (from ground)	mm	max 54	min 25	max 70	min 40
Tray height (from footrest)	mm	max 680	min 440	max 860	min 620

MOD.ORD.JUP.eng.rev.02

Italian Ministry of Health database REG.NO.2427909

Basic UDI 805050779005AM

